

PhD Thesis title: 'Evaluation of digital x-ray detectors for medical imaging applications'

Author: Anastasios C. Konstantinidis

Email: taskon25@yahoo.gr

Institution: University College London (UCL)

Supervisors: Prof. Robert Speller, Dr. Alessandro Olivo

Graduation Date: 28 July 2011

Available on line: <http://discovery.ucl.ac.uk/1322919/>

ABSTRACT:

Digital x-ray detectors are now the detector of choice in many X-ray examinations. They have been accepted into clinical practice over the past decade but there are still ongoing developments in their technology. Complementary metal oxide semiconductor (CMOS) active pixel sensors (APS) are a novel digital technology that offers advantages compared to some of the more established approaches (charge-coupled devices (CCD), thin film transistor arrays (TFT) and CMOS passive pixel sensors (PPS)). This thesis looks at the performance of these new sensors and attempts to identify their role in future medical imaging applications. Standard electro-optical and x-ray performance evaluations of two novel CMOS APS, namely the large Area Sensor (LAS) and Dexela CMOS x-ray detector, are presented. The evaluation was made in terms of the photon transfer curve (PTC), the modulation transfer function (MTF), the normalized noise power spectrum (NNPS) and the resultant detective quantum efficiency (DQE). Modifications were introduced to extend the standard methods to overcome technical limitations. The performance of these detectors was compared to three commercial systems (Remote RadEye HR (CMOS APS), Hamamatsu C9732DK (CMOS PPS) and Anrad SMAM (a-Se TFT)) at beam qualities (28 kV for mammography and 52 kV and 74 kV for general radiography) based on the IEC standards. Both the LAS and Dexela CMOS detectors demonstrate enhanced performance. The effect of the CMOS APS inherent nonlinearity on the x-ray performance was also evaluated. Finally, the measured performance parameters were used to simulate images for different mammographic imaging tasks in order to establish possible areas of application for the new sensors. Two software phantoms (one representing a 3-D breast and the other the CDMAM test tool) were used to simulate a range of mammographic conditions. The results show that both novel CMOS APS detectors offer high image quality compared to the commercial detector systems.

References to author publications that relate specifically to the dissertation:

- 1) **A. C. Konstantinidis**, M. B. Szafraniec, L. Rigon, G. Tromba, D. Dreossi, N. Sodini, P. F. Liaparinis, S. Naday, S. Gunn, A. McArthur, R. D. Speller and A. Olivo, "X-ray performance evaluation of the Dexela CMOS APS X-ray detector using

monochromatic synchrotron radiation in the mammographic energy range", IEEE Transactions on Nuclear Science, Volume 60, October 2013.

- 2) **A. C. Konstantinidis**, M. B. Szafraniec, R. D. Speller and A. Olivo, "The Dexela 2923 CMOS X-ray detector: A flat panel detector based on CMOS active pixel sensors for medical imaging applications", Nuclear Instruments and Methods in Physics Research A, Volume 689, October 2012, Pages 12-21.
- 3) **A. C. Konstantinidis**, A. Olivo and R. D. Speller, "Technical Note: Modification of the standard gain correction algorithm to compensate for the number of used reference flat frames in detector performance studies", Medical Physics, Volume 38, December 2011, Pages 6683-6687.
- 4) **A. C. Konstantinidis**, A. Olivo and R. D. Speller, "Technical Note: Further development of a resolution modification routine for the simulation of the modulation transfer function (MTF) of digital x-ray detectors", Medical Physics, Volume 38, November 2011, Pages 5916-5920.
- 5) **A. C. Konstantinidis**, A. Olivo, P. R. T. Munro, S. E. Bohndiek and R. D. Speller, "Optical characterisation of a CMOS active pixel sensor using periodic noise reduction techniques", Nuclear Instruments and Methods in Physics Research A, Volume 620, August 2010, Pages 549-556.
- 6) H. M. Zin, **A. C. Konstantinidis**, E. J. Harris, A. Olivo, S. E. Bohndiek, A. Clark, R. Turchetta, N. Allison, R. Speller and P. M. Evans, "Characterisation of regional variations in a stitched CMOS APS", Nuclear Instruments and Methods in Physics Research A, Volume 620, August 2010, Pages 540-548.
- 7) S. E. Bohndiek, A. Blue, J. Cabello, A. T. Clark, N. Guerrini, P. M. Evans, E. J. Harris, **A. C. Konstantinidis**, D. Maneuski, J. Osmond, V. O'Shea, R. D. Speller, R. Turchetta, K. Wells, H. Zin and N. Allinson, "Characterization and Testing of LAS: A Prototype 'Large Area Sensor' With Performance Characteristics Suitable for Medical Imaging Applications", IEEE Transactions on Nuclear Science, Volume 56, October 2009, Pages 2938-2946.